



Radiation Control, X-ray Unit
 625 North Robert Street
 P.O. Box 64975
 St. Paul, MN 55164-0975
 651-201-4545
www.health.state.mn.us/xray

EXAMPLE REPEAT ANALYSIS FORM

From: 01/01/03 To: 03/31/03

Facility: Hometown Dental

Cause	Number of Films	Percentage of Repeats
Black Films	2	5.8%
Dark Films	10	29%
Double Exposure	2	5.8%
Fog	3	8.8%
Incorrect Patient ID	2	5.8%
Light Films	4	11.7%
Patient Motion	11	32.3%
Total Film Used = 500		
Total Repeats	34	6.8%

CALCULATIONS: 34 total repeats/500 total film used = .068 or 6.8%
 2 total black films/34 total repeats = .058 or 5.8%