

Facility Inspection Checklist - Industrial

X-RAY UNIT

Documentation to Review

The inspection will include a review of all required documentation, including but not limited to:

- Alarm system entrance control test results, if applicable
- Calibration records for radiation survey instruments
- Delegation Agreement (ex: Radiation Safety Officer)
- Documentation from last inspection, if applicable
- Dosimetry records, if applicable
- Emergency procedures for x-ray equipment failures
- Instruction and training procedures and records
- Operating and emergency procedures
- Quality control records, if applicable
- Radiation Program Audit(s)
- Radiation Survey procedures and records
- Radiographer certification records, if applicable
- Shielding Plan and Radiation Survey documentation, if applicable
- Temporary job site records, if applicable
- X-ray equipment maintenance procedures and records
- X-ray equipment utilization log, if applicable
- Additional information may be required

An exit interview will be conducted to discuss the inspection preliminary findings.

Minnesota Department of Health
Radiation Control, X-ray Unit
625 Robert St N
PO Box 64975
St. Paul, MN 55164-0975
651-201-4545
health.xray@state.mn.us
www.health.state.mn.us/xray

01/13/20

To obtain this information in a different format, call: 651-201-4545.